

M8001-60024 Bedhanger for MP20/30

Installation Procedure - English

- 1 Remove the two screws securing the handle and remove the handle.



- 2 Using the same screws secure the bedhanger to the monitor.



CAUTION Make sure the bedhanger is tightly secured to the monitor (1.3 Nm torque) before hanging the monitor on a bed. Loose screws may result in the monitor separating from the bedhanger and falling down.

Printed in Germany 05/05

Part number M8001-9341A

4512 610 09851



PHILIPS

